

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130584

Supervisor/Operator Performing the Verification Procedure:

Name Supt. James R. R. Rea

ID# 3161

Date 04-23-09

A Agency Kotzebue Police Department

Phone # 907-442-3352

Instrument Location Kotzebue Regional Jail 258d rm#107 Fourth Avenue Kotzebue, AK 99752

Alco S/N X172931

Target Value .084

High Pressure 0650

B

Alco Test Values:

.084
1st Alco

.083
2nd Alco

Signature

Supt. James R. R. Rea

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


Nita J. Bolz

Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

Carolyn M. Noland

Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130584 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

[Empty box for taping the test strip]

ALASKA DEPARTMENT OF PUBLIC SAFETY
INSTRUMENT CALIBRATION

APRIL 23, 2005 ✓

OPERATOR'S NAME: ✓

REG. LICENSE NO. ✓

OPERATOR'S NUMBER: 3161

SUBJECT'S LAST NAME: ✓

VERIFICATION OF CAL

SUBJECT'S FIRST NAME (M): ✓

VERIFICATION OF CAL

O.L. # 1-100000

DEPT. AGENCY: OTZ1

TEST REPORT: 8480009

TEST TYPE: ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

DIAGNOSTIC CHECK

[Empty box for taping the test strip]

ALASKA DEPARTMENT OF PUBLIC SAFETY

INSTRUMENT CALIBRATION

APRIL 23, 2005 ✓

TIME 14:08

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/01

WEIGHTS: OKAY

DRINK CHARTER: 49c

WEIGHT TUBE: 41c

BAROMETER: 29.58

FLOW DETECTOR: OKAY

PUMP: OKAY

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STARTUP: OKAY

CALIBRATION: OKAY

PRINTER TEST

4/23/05 15:08:45 673711/01000000

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓

ALCO. TARGET VALUE: 0.04 ✓